#### REF: LMB-01

## **Instructions for use**



# **Kovacs Reagent (KOVACS)**

In vitro diagnostic medical aid.

Only for professional use.

**Purpose:** The product is designed for the detection of indole production in microorganisms.

Package contents: 500 ml solution in a bottle

Composition: N-amyl alcohol

hydrochloric acid

4- (dimethylamino) benzaldehyde

**Storage:** Store in a cool place at +2 to +8°C.

Usability: Use within 30 days after opening the primary packaging.

**Procedure:** Use the solution according to internal procedures.

**Information for users:** Do not use after the expiration date.

Dispose of used and unused products as hazardous waste in accordance with applicable regulations.

Warning:



H226 Flammable liquid

H302 Harmful if swallowed H315 Irritating to skin

H315 Irritating to skin H318 Causes serious e

H318 Causes serious eye damage.H335 May cause respiratory irritation

H336 May cause drowsiness and dizziness

P261 Avoid inhaling dust/smoke/gas/mist/vapor/aerosols. P280 Wear protective

gloves / eye protection / face protection.

P305 + P351 + P338 FOLLOWING EYE CONTACT: Gently rinse eyes with water for a few minutes. Remove contact lenses, if possible. Continue rinsing.

Instructions in case of a damaged packaging: In case of damage to the packaging, do not use and dispose of the product.

#### Quality control:

| Type of inspection | Inoculated culture            | Result                         |
|--------------------|-------------------------------|--------------------------------|
| positive check     | Proteus vulgaris CCM 1799     | Reagent color change to red    |
|                    | Escherichia coli CCM 3954     |                                |
| negative check     | Enterobacter cloacae CCM 1903 | No change in the reagent color |

### Manufacturer

EnviroLab s.r.o. Galvaniho 7 821 04 Bratislava Slovak Republic

## Place of business:

Obchodná ulica 7 946 51 Nesvady Slovak Republic

Last revision of the text: 11/2017